

2020-S-0004

Standard for Interpreting, Comparing and Reporting DNA Test Results Associated with Failed Controls and Contamination Events

*Human Forensic Biology Subcommittee
Biology Scientific Area Committee
Organization of Scientific Area Committees (OSAC) for Forensic Science*



*2020-S-0004 Standard for Interpreting,
Comparing and Reporting DNA Test Results Associated with
Failed Controls and Contamination Events*

Draft OSAC Proposed Standard

2020-S-0004 Standard for Interpreting, Comparing and Reporting DNA Test Results Associated with Failed Controls and Contamination Events

Prepared by
Human Forensic Biology Subcommittee
Version: 1.0
December 2020

Disclaimer:

This OSAC Proposed Standard was written by the Human Forensic Biology Subcommittee/Biology Scientific Area Committee of the Organization of Scientific Area Committees (OSAC) for Forensic Science following a process that includes an [open comment period](#). This Proposed Standard will be submitted to a standards developing organization and is subject to change.

There may be references in an OSAC Proposed Standard to other publications under development by OSAC. The information in the Proposed Standard, and underlying concepts and methodologies, may be used by the forensic-science community before the completion of such companion publications.

Any identification of commercial equipment, instruments, or materials in the Proposed Standard is not a recommendation or endorsement by the U.S. Government and does not imply that the equipment, instruments, or materials are necessarily the best available for the purpose.

To be placed on the OSAC Registry, certain types of standards first must be reviewed by a Scientific and Technical Review Panel (STRP). The STRP process is vital to OSAC's mission of generating and recognizing scientifically sound standards for producing and interpreting forensic



*2020-S-0004 Standard for Interpreting,
Comparing and Reporting DNA Test Results Associated with
Failed Controls and Contamination Events*

science results. The STRP shall provide critical and knowledgeable reviews of draft standards or of proposed revisions of standards previously published by standards developing organizations (SDOs) to ensure that the published methods that practitioners employ are scientifically valid, and the resulting claims are trustworthy.

The STRP panel will consist of an independent and diverse panel, including subject matter experts, human factors scientists, quality assurance personnel, and legal experts, which will be tasked with evaluating the proposed standard based on a comprehensive list of science-based criteria.

For more information about this important process, please visit our website at:

<https://www.nist.gov/topics/organization-scientific-area-committees-forensic-science/scientific-technical-review-panels>

1 Foreword

2 Controls are routinely incorporated during DNA testing of forensic and reference
3 samples in forensic DNA testing laboratories. If all controls generate the expected
4 results, there is a high level of confidence in the profile data from the associated sample
5 set. A control may fail for various reasons. In addition, the profile data may indicate a
6 handling error or the presence of contaminating DNA. Retesting the forensic sample
7 prior to the step in which the problem was identified may be performed; however there
8 are circumstances where this may not be feasible or necessary. Reasons for not
9 conducting retesting include, but are not limited to, the sample was consumed during
10 the initial analysis, additional testing would exhaust the remaining portion of the sample
11 or DNA extract eliminating the possibility of future testing, or the associated profile(s)
12 would not be suitable for comparison even if the controls produced the expected results.

13 There are scenarios where it may be possible to interpret, compare, and report data
14 with some level of confidence, even if the data are associated with the failure of a
15 control or a contamination event (of a sample or control). Evaluation and reporting of
16 the possibly compromised data may provide critical and valid information to support
17 the investigation of a criminal case, for example excluding a person of interest. To the
18 extent determination of contamination may be influenced by judgmental bias, persons
19 making that determination should be shielded from irrelevant information.

20 It is intended that this standard be used in conjunction with the laboratory's
21 documented quality assurance program. This would ensure that proper evaluations,
22 root cause analyses, risk assessments, and corrective actions, when necessary, have
23 been performed and appropriately documented for each instance of a failed control or
24 contamination event that occurs in the laboratory. It is also intended that the laboratory
25 perform the requirements in this standard using documented protocols for data
26 interpretation, comparison and reporting with appropriate accompanying validation
27 and protocol verification studies along with the strong reliance on other available
28 standards for forensic DNA testing (e.g., FBI Quality Assurance Standards for DNA
29 Testing Laboratories, ANSI/ASB Standards 18, 20, 40, 136 and 139 and OSAC Best
30 Practices Recommendations for the Management and Use of Quality Assurance DNA
31 Elimination Databases in Forensic DNA Analysis; see Bibliography). This document is
32 not intended to support the reporting of data associated with failed controls and/or
33 contamination events without the associated prerequisite for thorough evaluation of the
34 possible cause and impact of the events on the data obtained.

35 The draft of this standard was developed by the Biological Data Interpretation and
36 Reporting Subcommittee of the Organization of Scientific Area Committees for Forensic
37 Science.

38 All hyperlinks and web addresses shown in this document are current as of the
39 publication date of this standard.

40 **Keywords:** *contamination, failed control, reporting DNA results, DNA interpretation*



41	Table of Contents	
42	1 Scope	
43		
44	2 Normative References	
45		
46	3 Terms and Definitions	
47		
48	4 Requirements	
49	Annex A (informative) Supplemental Information – Foundational Principles	3
50	Annex B (informative) Supplemental Information – Examples	5
51	Annex C (informative) Bibliography	

52 **Standard for Interpreting, Comparing and Reporting DNA test Results**
53 **Associated with Failed Controls and Contamination Events**

54 **1 Scope**

55 This standard provides requirements for the interpretation, comparison, and reporting of DNA data
56 associated with control failures or contamination where re-testing is not performed. DNA data
57 associated with a failed control or a contamination event may still be scientifically valid and may
58 be relevant to an investigation. These standards may be applied to any type of forensic DNA testing
59 technology and methodology when conducted in an accredited forensic laboratory.
60

61 **2 Normative References**

62 There are no normative reference documents. Annex C, Bibliography, contains informative
63 references.

64 **3 Terms and Definitions**

65 For purposes of this document, the following definitions apply.

66 **3.1**

67 **comparison**

68 The process of examining two or more DNA data sets to assess the degree of similarity or
69 difference.
70

71 **3.2**

72 **contamination**

73 Exogenous DNA or other biological material in a DNA sample, PCR reaction, or item of
74 evidence; the exogenous DNA or biological material could be present before the sample is
75 collected, or introduced during collection or testing of the sample.

76 **3.3**

77 **failed control**

78 A positive control (3.7) or negative control (3.6) that produces an unexpected result.
79

80 **3.4**

81 **forensic sample**

82 A biological sample originating from and associated with evidence from a crime scene. A
83 sample associated with evidence from a crime scene may include a sample that has been
84 carried away from the crime scene.
85

86 **3.5**

87 **interpretation**

88 The process of evaluating DNA data for purposes including, but not limited to, defining
89 assumptions related to mixtures and single source profiles, distinguishing between alleles

90 and artifacts, assessing the possibility of degradation, inhibition, and stochastic effects, and
91 determining whether the data are suitable for comparison.

92

93 **3.6**

94 **negative control**

95 An analytical control that consists of the reagents used in various stages of testing
96 without the introduction of sample; no results are expected from a negative control. For
97 DNA testing, negative controls include, but are not limited to, extraction blanks/reagent
98 blanks and amplification blanks. A negative control in DNA testing is used to detect
99 contamination introduced into the assay during the testing process via reagents,
100 disposables or handling errors (which may impact the results observed from samples
101 tested at the same time).

102

103 **3.7**

104 **positive control**

105 An analytical control sample that is used to determine if a test performed properly. This
106 control consists of the test reagents and a known sample that will provide an expected
107 positive response with the test. For DNA testing, positive controls may include, but are
108 not limited to, extraction positive controls and positive amplification controls.

109

110 **3.8**

111 **reference sample**

112 Biological material obtained from a known individual and collected for purposes of
113 comparison to evidentiary samples.

114

115 **3.9**

116 **suitable for interpretation**

117 Data deemed appropriate for interpretation (3.5) based on the laboratory's validation
118 studies and documented and verified interpretation protocol.

119

120 **3.10**

121 **unsuitable for comparison**

122 Data that cannot be used for comparisons for reasons including, but not limited to, poor or
123 limited data quality, mixture complexity, or a failure to meet quality assurance
124 requirements. This decision is based on the laboratory's validation studies and
125 documented and verified interpretation and comparison protocol.

126

127 **4 Requirements**

128 **4.1** The laboratory protocol shall define what constitutes:

129

130 **4.1.1** Contamination in a negative control

131

132 **4.1.2** Contamination in a positive control

133

134 **4.1.3** Contamination in forensic or reference sample DNA test results

135

136 **4.1.4** A failed positive control

137
138
139
140
141
142
143
144
145
146
147
148
149
150
151
152
153
154
155
156
157
158
159
160
161
162
163
164
165
166
167
168
169
170
171
172
173
174
175
176

4.1.5A failed negative control

4.2 The laboratory shall perform and document the assessment of the integrity of the associated DNA test results to determine the impact of the failed control or contamination. The assessment shall be based on scientifically valid principles in DNA analysis and include, as appropriate, a determination of the possible cause and effect of the failed control or contamination, and an assessment of the risks associated with moving forward with data interpretation vs. those associated with re-testing.

4.2.1 If the DNA test results are determined to be suitable for interpretation within the constraints of the laboratory's internal validation studies and documented interpretation and comparison protocols and the laboratory does not retest, the laboratory shall perform and report the interpretation and comparison(s) with applicable statistical analysis.

4.2.2 If the DNA test results are determined to be compromised to the extent of being unsuitable for interpretation and retesting is not conducted, the results shall be reported as not suitable for interpretation according to laboratory policy.

NOTE If the DNA test results are determined to be compromised to the extent of being unsuitable for interpretation and retesting is conducted, it may be necessary to report results, interpretations and comparisons, as appropriate, from both the original and second tests.

4.3 When reporting interpretations and comparisons impacted by a failed control or contamination event, the report shall identify the associated DNA test results and a description of the nature of the event.

4.4 The laboratory shall have a written protocol for the release of identifying information for the source of the contamination.

4.5 The case record for each sample associated with a failed control or contamination event must include documentation of the following for the affected sample(s), as applicable:

4.5.1 The forensic sample, reference, or control DNA test result that failed or was contaminated.

4.5.2 The likely or known source of contamination.

NOTE The source may be identified by name, employment position or other descriptor as permitted by law and agency policies.

177

178

179 **4.5.3** The likely or known cause of the failed control or contamination.

180

181 **4.5.4** The impact of the failed control or contaminant on the integrity of the DNA test results.

182

183 **4.5.5** The determination of whether an affected DNA test result is suitable, or unsuitable, for
184 interpretation.

185

186

187

188

189

190

191

192

193

194

195

196

197

198

199

200

201

202

203

204

205

206

Annex A

207

(informative)

208

Supplemental Information – Foundational Principles

209 When polymerase chain reaction (PCR) testing was introduced into crime laboratories in the early
210 1990s, many of the initial issues encountered by researchers using the highly sensitive PCR testing
211 methods had been recognized. As a result, standard procedures for preventing contamination
212 along with quality control and assurance measures were established in forensic DNA testing
213 laboratories. Even with these critical measures in place, occasional problems occur during DNA
214 testing. Forensic DNA testing and databasing laboratories typically have a number of processes in
215 place for monitoring and evaluating the integrity of the DNA testing results obtained from samples
216 received and processed by the laboratory. When the laboratory identifies instances where the DNA
217 test results may be compromised, the laboratory follows required procedures for establishing the
218 likely cause of the event and for assessing its impact on the data obtained. This impact assessment
219 step is critical in that the DNA test results may still be valid and further interpretation may provide
220 valuable information, such as exculpatory evidence.

221

222 When performing PCR testing, forensic DNA testing laboratories are required to have a positive
223 amplification control associated with each set of DNA extracts amplified together^[3]. This control
224 monitors the DNA testing process performed through all steps commencing at the amplification
225 step. Some laboratories require an additional positive control to be associated with the DNA
226 extraction batch that then follows the samples through the entire DNA testing process. At the end
227 of testing, DNA test results from the positive control(s) should be consistent with the expected
228 reference single source profile(s). The presence of the correct DNA test results in the positive
229 control indicates the testing process(es) monitored by the control(s) performed correctly.

230

231 There are several possible causes for a positive control failure (as defined by the laboratory),
232 including a technical issue (e.g., problem with an instrument or reagent that precludes the test from
233 working correctly). Similar issues may have occurred with the associated samples. When it is not
234 possible to use the results due to a concern of accuracy, then retesting starting from a point before
235 the instrument or reagent issue is necessary to generate test results that can be reliably
236 interpreted, compared and reported. If retesting is not possible and the integrity of the DNA test
237 results cannot be confirmed, the results may be reported as “insufficient for comparison” or
238 “inconclusive” due to the control failure.

239

240 In some cases, the positive control failure may be determined to be specific to only that sample,
241 with the other DNA test results processed with the control seemingly unaffected. This may occur,
242 for example, if DNA or reagents were inadvertently not added to the control but added correctly for
243 the other DNA extracts. In this case, it may be possible to verify that the other results associated
244 with the failed control can be interpreted, compared and reported after fulfilling the requirements
245 of this standard without retesting all of the samples involved.

246

247 In addition, when performing PCR testing, forensic DNA testing laboratories are required to have
248 two negative controls associated with each set of DNA samples tested^[3]. One negative control,
249 typically referred to as a reagent blank or extraction blank control, is started with each set or batch
250 of samples extracted together; the second negative control is the negative amplification control
251 started at the amplification step for each set of samples undergoing amplification together. These
252 two negative controls are processed throughout each step of the DNA test alongside the associated
253 samples. These two controls consist of all reagents, solutions, consumable materials, etc. used during
254 the DNA testing process, and it is expected that the negative controls meet the laboratory’s definition

255 for suitable performance when evaluated at the end of the testing. When contamination is identified,
256 the laboratory is responsible for evaluating the likely biological source of the contamination and
257 assessing when and how the event most likely occurred.
258

259 In some situations, the contaminating DNA is only detected in a negative control with no apparent
260 presence in or effect on any of the other samples tested. This single contamination event may be
261 due to any number of reasons where DNA could be introduced only into a single sample, for
262 example, its presence in or on a consumable material used in the laboratory during testing (e.g.,
263 pipet tip, tube). In other situations, the contaminating DNA may be detected in the profiles from
264 other samples tested along with the control(s) but be present at such a low level that it has
265 minimal to no impact on the quality of the DNA test results obtained from the tested sample (e.g.,
266 DNA profile from a high quality single source or two person mixed DNA profile with a very low
267 level minor component consistent with the profile in the negative control and possibly other
268 samples). In these cases, the DNA test results may be reasonably determined, interpreted and used
269 for comparison according to established laboratory protocols in accordance with the requirements
270 listed in this document.
271

272 DNA contamination may also be present in one or more of the forensic or reference samples. Many
273 laboratories have internal DNA databases comprised of DNA data from laboratory or other
274 personnel who may routinely come into contact with samples or be present in the environment
275 where forensic samples are handled or processed (e.g., law enforcement, evidence technicians,
276 crime scene investigators, maintenance staff). These DNA databases may be used as a screening
277 mechanism for the detection of possible DNA contamination events (also see Best Practice
278 Recommendations for the Management and Use of Quality Assurance DNA Elimination Databases
279 in Forensic DNA Analysis). Similarly, some laboratories compare the data obtained within certain
280 test batches to screen for possible contamination events that may have occurred between DNA
281 extracts processed concomitantly. During these evaluations, the source of the contaminating DNA
282 may be identified. In this situation, it may be possible to evaluate the DNA test results even in the
283 presence of contaminating DNA from a known individual, similar to the interpretation steps used
284 to evaluate mixed DNA test results when a known contributor to a DNA mixture is assumed. The
285 use of an assumed contributor in the interpretation and comparison of the data should be
286 reported according to the laboratory's protocol and best practice recommendations for reporting
287 evaluations performed using assumed contributors.
288

289 Additional standards and best practice recommendations are referenced in the Bibliography
290 that may be used in conjunction with this standard and provide additional useful information

291
292
293

Annex B **(informative)**

294

Supplemental Information – Examples

295 The following examples describe different scenarios where samples are associated with a failed
296 control or contamination event with some possible outcomes responsive to the requirements of
297 this standard:

- 298
- 299 1. No results were obtained for the amplification positive control and the associated forensic
300 samples provided partial or full profiles that corresponded logically to their respective
301 quantitation results. The laboratory investigates and determines the most likely cause was
302 that the analyst did not add the known DNA to the amplification positive control sample. A
303 surrogate control (in this example, a positive control from the previous day's run on the
304 same electrophoresis instrument), was used to confirm that the allele calling was
305 performed correctly by the software and the profiles were interpreted and used for
306 comparison purposes. The issue and resolution were documented in the case record and
307 the results were reported per the laboratory protocol since the results were not directly
308 impacted by the failed control.
 - 309 2. The DNA profile of a member of the laboratory was detected as a minor component of a
310 two person mixture profile detected from a forensic sample. The laboratory staff member
311 was the individual who performed the latent print examination on the sample prior to the
312 DNA testing. The DNA profile was interpreted and used for comparison under the
313 assumption that the laboratory staff member was one of the contributors to the mixture.
314 Since the interpretation was directly impacted, the contamination event was described in
315 the report.
 - 316 3. A low level DNA profile was detected in the extraction reagent blank that was consistent
317 with the low level DNA profile detected from the forensic sample. The forensic sample and
318 DNA extract were consumed during testing. Investigation could not determine the cause of
319 the contamination event (e.g., whether cross contamination occurred or whether the
320 reagents themselves were contaminated). The results for the forensic sample were
321 reported as not suitable for comparison purposes. Since the contamination event directly
322 impacted the interpretation of the profile from the forensic sample, the contamination
323 event was described in the report.
 - 324 4. The DNA profile from the forensic sample associated with a failed positive control
325 demonstrated the presence of a mixture of at least six individuals. The assessment of the
326 impact of the failed positive control determined that the interpretation of the forensic
327 sample profile was not affected since the laboratory's protocol does not permit the
328 interpretation of mixtures of greater than four individuals. No retesting was performed; the
329 forensic sample profile was reported as not suitable for comparison purposes due to the
330 high number of contributors.
 - 331 5. The DNA profile of the working DNA analyst was detected in the epithelial cell fraction of a
332 sexual assault kit sample and there was no indication of contamination of the sperm
333 fraction. Because the remaining contributor profile in the epithelial cell fraction was
334 consistent with the complainant, retesting was not performed. Results from both the
335 epithelial cell fraction and sperm cell fraction were interpreted, used for comparison and
336 reported. Since the contamination event directly impacted the interpretation of the
337 forensic sample profile, the contamination event was described in the report
 - 338 6. The DNA profile of the technician who performed amplification set up was detected in

339 the negative amplification control. A review of the associated samples shows that they
340 were not impacted by the contamination and no retesting was performed. The issue and
341 resolution were documented in the case record and the results were reported per the
342 laboratory protocol since the results were not directly impacted by the contamination.

343 7. The DNA profiles from an amplification plate show a low-level contaminant throughout,
344 indicating that there may have been contamination of the amplification master mix.
345 Because of the way the contaminant presents, the associated forensic sample profiles
346 were determined to be unsuitable for comparison. The DNA amplified includes the
347 consumed extract of a single swab (also consumed) from the neckline of a shirt. The
348 neckline of the shirt is resampled by taking and consuming a second swab, and an
349 interpretable profile is obtained. The laboratory report should address both the first and
350 second sampling of the neckline of the shirt.

351
352
353
354
355
356
357
358
359
360
361
362
363
364
365
366
367
368
369
370
371
372
373
374
375
376
377
378
379
380
381
382
383
384
385
386
387
388
389
390
391
392
393
394
395
396

Annex C
(informative)

Bibliography

- 1] Forensic Science Regulator. Forensic Science Providers: Codes of Practice and Conduct. 2014. Available at <https://www.gov.uk/government/collections/forensic-science-providers-codes-of-practice-and-conduct>
 - 2] SWGDAM Contamination Prevention and Detection Guidelines for Forensic DNA Laboratories. Available at www.swgdam.org
 - 3] Federal Bureau of Investigation. Quality Assurance Standards for Forensic DNA Testing Laboratories. Current version available at www.swgdam.org
 - 4] ISO/IEC 17025 Testing and Calibration Laboratories. Current version available at <https://www.iso.org/ISO-IEC-17025-testing-and-calibration-laboratories.html>
- ANSI/ASB and/or OSAC Standards and Best Practices Applicable to this Standard
- 5] OSAC Proposed Standard 2020-N-007, Best Practice Recommendations for the Management and Use of Quality Assurance DNA Elimination Databases in Forensic DNA Analysis; Current version available at [2020-N-007, Best Practice Recommendations for the Management and Use of Quality Assurance DNA Elimination Databases in Forensic DNA Analysis](#). (accessed February 22, 2021) [Update link]
 - 6] ANSI/ASB Standard 136, Forensic Laboratory Standards for Prevention, Monitoring, and Mitigation of DNA Contamination, Ed. #1 (currently in review at ASB; has been out for public comments)
 - 7] ANSI/ASB Standard 018, Validation Standards for Probabilistic Genotyping Systems First Edition, 2020 http://www.asbstandardsboard.org/wp-content/uploads/2020/07/018_Std_e1.pdf (accessed October 2020)
 - 8] ANSI/ASB Standard 020, Standard for Validation Studies of DNA Mixtures, and Development and Verification of a Laboratory’s Mixture Interpretation Protocol, First Edition, 2018 https://asb.aafs.org/wp-content/uploads/2018/09/020_Std_e1.pdf (accessed December 2020)
 - 9] ANSI/ASB Standard 040, Standard for Forensic DNA Interpretation and Comparison Protocols, First Edition, 2019 http://www.asbstandardsboard.org/wp-content/uploads/2019/10/Std_040_e1.pdf (accessed December 2020)
 - 10] ANSI/ASB Standard 139, Reporting DNA Conclusions (currently in review at ASB; has been out for public comments)