

Reference Bullets and Cartridge Cases for Automated Ballistics Inspection Systems

T.V. Vorburger (tvtv@nist.gov), W. Chu, A. Zheng, T.B. Renegar,
J. Yen, J.F. Song, J.A. Soons, R.M. Thompson
National Institute of Standards and Technology (NIST)
Gaithersburg, MD

M.G. Ols
Bureau of Alcohol, Tobacco, Firearms, and Explosives (ATF)
Ammendale, MD



AAFS 66th Annual Scientific Mtg.
Seattle, WA
February 21, 2014



Contents

- Introduction to SRM Bullets and Cartridge Cases
- Topography and Microscopy Images of SRM Cartridge Cases
- Reproducibility
- Traceability and User Acquisition
- Availability

Funding Provided by

- National Institute of Justice
NIST Law Enforcement Standards Office
NIST Standard Reference Materials Program

Note:


Certain commercial equipment may be identified in this presentation in order to specify an experimental procedure. This does not imply recommendation or endorsement by NIST, nor does it imply that the equipment are the best available for the purpose.

SRM 2460 Standard Bullet and SRM 2461 Standard Cartridge Case

- For demonstrating consistency in image acquisitions of bullets and cartridge cases from place to place
- Aimed at automated ballistic inspection systems, such as those in the National Integrated Ballistic Information Network (NIBIN)
- Requires high degree of similarity from one SRM unit to another





SRM 2460 Standard Bullets manufactured by numerically controlled diamond turning



40 SRM Bullets were produced and distributed.

SRM 2461 Standard Cartridge Cases manufactured by commercial electroforming


- Negatives are made from a master cartridge case
- Positive replicas are then made from the negatives



SRM Cartridge Cases became available in July 2012.

Image Acquisition

Confocal microscopy used to acquire 3D **surface topography images** of SRM Cartridge Cases




NIBIN stations at ATF's National Laboratory Center used to acquire **optical microscope images** of SRM Cartridge Cases.




Photo: www.forensictechnology.com


Topography Images for a SRM Cartridge Case



Ejector mark



Firing Pin




Breech Face

These images should be essentially **the same** for all units of the SRM.


NIBIN *Golden Images* for the Standard Cartridge Case Using Optical Microscopy at ATF National Lab. Ctr.



Ejector mark



Firing Pin




Breech Face

These images should be essentially **the same** for all units of the SRM.

Contents

- Introduction to SRM Bullets and Cartridge Cases
- Topography and Microscopy Images of SRMs
- **Reproducibility**
- Traceability and User Acquisition
- Availability

SRM 2461 Analysis – Breech Face



Processing steps:

- Trimming
- Removal of outlier data
- Form & noise filtering
- Image registration
- Correlation calculation



National Institute of Standards & Technology
Certificate

Standard Reference Material® 2460

Standard Bullet
Serial No.: SAMPLE

Standard Reference Material (SRM) 2460 is a bullet signature standard comprising bullet profile signatures of six lead engraved areas (LEAs) from fired bullets. This SRM is intended primarily for use as a check standard for crime laboratories to help verify that the computerized optical equipment for bullet imaging and profiling is operating properly. A unit of SRM 2460 consists of an SRM standard bullet that is mounted on a blue knob (see Figure 1).

A Virtual-Physical Bullet Signature Standard. The SRM 2460 physical bullet signature standard is derived from a virtual standard, as shown in Figure 2. It is a set of six digitized bullet profile signatures that provided the information for machining the bullet signatures on the physical standard. The SRM 2460 standard bullets [1,2]. The virtual standard also provides the reference profiles for comparison measurements of these bullet signatures [3].

Certified Cross Correlation Function Maximum CCF_{max} and Signature Difference D . The certified values for cross correlation function maximum CCF_{max} and signature difference D are based on results obtained from profile comparisons between the six profile signatures on the SRM 2460 standard bullet and those of the virtual bullet signature standard. For an ideal match between the bullet signature and the virtual standard, CCF_{max} is equal to 1 and D is equal to 0. Forty SRM 2460 standard bullets, with serial numbers 521 001 to 521 040, were measured and compared with the virtual standard. The measurement results were statistically analyzed. The values of six cross correlation function maxima CCF_{max} and signature differences D for the six bullet signatures were averaged for each SRM 2460 bullet. These averages were designated CCF_{max} and D . For the 40 SRM standard bullets, the collective lower limit for CCF_{max} and upper limit for D , with a 95 % confidence level ($n = 95$ %) [4] are reported in Table 1. A NIST certified value is a value for which NIST has the highest confidence in its accuracy as that all known or suspected sources of bias have been investigated or accounted for by NIST [5]. The current distribution of SRM 2460 consists of 35 bullets. These bullets were checked and rechecked periodically. Five of the 35 bullets were measured again to show that the etching process produced no discernible change in the surface topography measured by the stylus instrument.

Expiration of Certification: The certification of SRM 2460 is expected to be valid within the measurement uncertainties specified, until **30 June 2014**, provided the SRM is handled, stored, and used in accordance with the instructions given in this certificate (see "Notice and Warning to Users"). However, the certification is nullified if an inspected area that is damaged, contaminated, or modified. NIST reserves the right to withdraw, suspend, or extend this certification as appropriate.

Maintenance of SRM Certification: NIST will monitor this SRM over the period of its certification. If substantive surface changes occur that affect the certification before the expiration of this SRM, NIST will notify the purchaser. Repurchase (see attached sheet) will facilitate notification.

The coordination of SRM 2460 production and the technical measurements leading to the certification of this SRM were performed by J. Song and T. Vorburger of the NIST Precision Engineering Division, and S. Ballou of the NIST Office of Law Enforcement Standards (OLEES).

Theodore Deion, Acting Division Chief
Precision Engineering Division
Robert Waters, Jr., Chief
Measurement Services Division

Page 1 of 8

Gaithersburg, MD 20899
Certificate Issue Date: 30 October 2002
SRM 2460



National Institute of Standards & Technology
Certificate

Standard Reference Material® 2461

Standard Cartridge Case

This Standard Reference Material (SRM) is intended primarily for use as a check standard for crime laboratories, first to help verify that the computerized optical equipment for cartridge case imaging acquisition and correlation is operating properly, second, to establish ballistics measurement traceability and quality assurance, and third, to facilitate laboratory assessments and accreditation [1-3]. A unit of SRM 2461 consists of a circular electro-formed of a brass face impression, a firing pin impression, and an ejector mark. The electro-formed plate is cemented to a brass cylinder holder (see Figure 1) so that the assembly resembles a real fired cartridge case.

Certified Areal Cross Correlation Function Maximum ($ACCF_{max}$) and Signature Difference (D). $ACCF_{max}$ and D are the two properties of the surface topography used to characterize the similarity of the cartridge case surfaces and obtain the certified values in SRM 2461 [1]. A NIST certified value is a value for which NIST has the highest confidence in its accuracy in that all known or suspected sources of bias have been investigated and taken into account [4]. The certified values are obtained from statistical correlations between the surface topography of the brass face firing pin, and ejector mark impressions of 137 SRM 2461 standard cartridge cases and those of the reference standard to determine their degree of similarity. When two correlated cartridge cases and those of the same type by print, the lower limit of D equals 0 %, and the upper limit of $ACCF_{max}$ equals 100 %. The certified values for the degree of similarity in Table 1 are reported as one-sided intervals, each with 95 % confidence [4].

Table 1. Certified Areal Cross Correlation Function Maximum ($ACCF_{max}$) and Signature Difference (D).

	$ACCF_{max}^{(H)}$ (%)	$D^{(L)}$ (%)
Brass Face	-94.3	-11.2
Firing Pin	-98.0	-4.0
Ejector Mark	-93.7	-12.2

^(H) The one-sided interval with 95 % confidence represents the measurement uncertainty in similarity in SRM 2461. The measurement uncertainty is the range covered from the lower or upper limits to perfect similarity or zero difference respectively. Two surfaces cannot have a similarity better than perfect ($ACCF_{max} = 100$ %) or a difference less than nothing ($D = 0$).

Expiration of Certification: The certification of SRM 2461 is valid, within the measurement uncertainty specified, until **30 September 2011**, provided the SRM is handled and stored in accordance with the instructions given in this certificate (see "Notice and Warning to Users"). The certification is nullified if the SRM is damaged, contaminated, or otherwise modified.

Maintenance of SRM Certification: NIST will monitor this SRM over the period of its certification. If substantive surface changes occur that affect the certification before the expiration of this SRM, NIST will notify the purchaser. Repurchase (see attached sheet) will facilitate notification.

The coordination of the production and the technical measurements leading to the certification of SRM 2461 were performed by A. Zhang, T.B. Renner, S. Mostafa-Dalga, C. Jmicki, J. Villanova, J. Song, T.V. Vorburger, and R. Silver of the NIST Semiconductor and Dimensional Metrology Division and E.M. Thompson of the NIST Law Enforcement Standards Office (OLEES).


Gaithersburg, MD 20899
Certificate Issue Date: 22 June 2002
SRM 2461

David G. Seiler, Chief
Semiconductor and Dimensional Metrology Division
Robert L. Waters, Jr., Chief
Measurement Services Division
Page 1 of 11

NIST Statisticians participated in the analysis of the uncertainty estimates and control values, and preparation of the Certificate.

Both SRMs and their Certificates include User Guides as appendices

Two Traceability Chains





SRM 2461 Documentation Includes a User Guide mainly for NIBIN Users

APPENDIX

USER GUIDE FOR NIST SRM 2461 STANDARD CARTRIDGE CASES

A Draft Data Acquisition Procedure for the Integrated Ballistics Identification System (IBIS)

1. The NIST SRM standard cartridge case is intended for quality testing of automated ballistics signature acquisitions. It is recommended that an acquisition of the standard cartridge cases be entered into IBIS once a month to verify the proper operation of the system. However, you should refer to your own laboratory's policy and procedure guidelines (PPG's) for the frequency of acquisition. The cartridge case should also be entered into the system during each software and hardware upgrade as well as after any scheduled or unscheduled maintenance. All entries and results should be documented.



Example of Correlation Results for BrassTrax Acquisitions of SRM 2461

Correlation Requests							
Case Number	Exhibit Number	Caliber	Site Name	Type	Correlation Server	Status	Modified By
C11-26074	2A	45 Auto	IL-CHICSP6	Automatic	Default	Correlation done	CMDAIVSON
C11-26074	1A	9 mm Parabellum	IL-CHICSP6	Automatic	Default	Correlation done	CMDAIVSON
C10-48660	2-3	9 mm Luger	IL-CHICSP4	Automatic	Default	Correlation done	DMFRATT
C10-48660	2-2	9 mm Luger	IL-CHICSP4	Automatic	Default	Correlation done	DMFRATT
C10-48660	1-7	40 Smith & Wesson	IL-CHICSP4	Automatic	Default	Correlation done	DMFRATT
C10-39956	1-12	40 Smith & Wesson	IL-CHICSP4	Automatic	Default	Correlation done	KMMURRAY
C11-46651					Default	Correlation done	TLBRUBAKER
OLS/2011					Svr-Reg 6	Correlation done	MGOLS
C10-41163					Default	Previously Viewed	KLHUNTER
C11-10493					Default	Correlation done	MANALLY
C10-32780					Default	Correlation done	JAHANNA
C10-32700					Default	Correlation done	JAHANNA


Cartridge Case Correlation Results							
Case Number	Exhibit Number	Caliber	Site Name	Scores			
				Breach Face	Firing Pin	Ejector Mark	
TRAX GOLDEN CC IMAGE	TRAX GOLDEN 10/26/10	9 mm Parabellum	MD-WATFBR1	372	344	2361	
NIST GOLDEN IMAGE	GOLDEN IMAGE 5/16/08	9 mm Parabellum	MD-WAATF2	318	337	868	
COLS-87-04	EX-04000271-1C	9 mm Luger	OH-COLUMPD	24	49	N/A	
T05-764	2	9 mm Parabellum	VA-NORFDFS	24	71	N/A	
CANT02-83133DF	83133C	9 mm Parabellum	OH-STARKCO	23	32	N/A	
COLS-6851-08	EX-08018361-1C	9 mm Parabellum	OH-COLUMPD	23	20	N/A	
T08-4293	10	9 mm Parabellum	VA-NORFDFS	23	47	N/A	
T06-13202	3A	9 mm Parabellum	VA-NORFDFS	23	40	N/A	
03F05157	Q1JY	9 mm Luger	MD-BALTPD2	23	52	N/A	
COLS-8434-03	EX-03027195-1C	9 mm Luger	OH-COLUMPD	22	28	N/A	
09-097-1341	QC-1	9 mm Parabellum	MD-BALTCO	22	41	N/A	
30-09-25611	TF-1-C	9 mm Luger	DE-WILMPD2	22	11	N/A	

**Use "TRAX GOLDEN CC IMAGE"
for control chart scores**

Excel Sheet available from NIST with Control Chart for Tests of SRM Cartridge Cases

Table 2. A Control Chart for Breech Face (BF) Scores of SRM Cartridge Case (Sheet 3 "BF")

Date	BF Score	Dynamic CL	Fixed CL	File Name	Notes
3/23/2010	376		214	SRM2461-018-23MAR10A	
4/23/2010	403	358	214	to be added	
5/23/2010	368	352	214		
6/23/2010	422	351	214		
7/23/2010	357	341	214		
8/23/2010	398	347	214		
9/23/2010	447	344	214		
10/23/2010	382	345	214		
11/23/2010	363	342	214		
12/23/2010	427	344	214		
1/23/2011	414	348	214		
2/23/2011	379	348	214		
			214		
			214		



Proposed Fixed Control Limits for Acquisitions of SRM 2461 (and 2460) Established by the National Ballistics Imaging Comparison Project Involving about 30 NIBIN Participants and their Systems


	Mean	Std. Dev.	95 % Control Limit
Breech Face	329	53	243
Firing Pin	344	44	273
Ejector Mark	1476	562	552

Availability of SRM 2461

- Available from the NIST Standard Reference Materials Program (www.nist.gov/srm) since July 2012
- 37 units distributed
- 100 units available
- The certification of **SRM 2461** is valid until **30 September 2021**, with proper handling.

SRM 2460 Bullets not presently available

- Currently developing a polymer replication method for restocking



- See: Today, 10:40 a.m. - 11:00 a.m., Paper A170
An Improved Vacuum Casting Method for the Replication of Reference Bullets
Thomas B. Renegar et al.